

Test Report

Report No. RLSZD000842980004

Page 1 of 4

Applicant OFC CHENGDU APPLIED SCIENCE & TECHNOLOGY GROUP CORP.

Address CHENGDU HI-TECH ZONE, SICHUAN CHINA

Report on the submitted sample(s) said to be

Sample Name ALUMINIUM FLUX F105 SERIES

Sample Description Transparent liquid

Part No. F105 SERIES

Item No. 20110225041

Supplier OFC

Manufacturer OFC

Sample Received Date Mar. 17, 2011

Testing Period Mar. 17, 2011 to Mar. 18, 2011

Test Requested As specified by client, to determine the Lead(Pb), Cadmium(Cd), Mercury(Hg), Hexavalent Chromium(Cr(VI)), Chlorine(Cl), Bromine(Br) content in the submitted sample.

Test Method Please refer to the following page(s).

Test Result(s) Please refer to the following page(s).

Tested by



Approved by

Inspected by

Vargas

Date

Mar. 18, 2011

No. 21208394

Test Report

Report No. RLSZD000842980004

Page 2 of 4

Test Method

Test Item(s)	Test Method	Measured Equipment(s)	MDL
Lead(Pb)	IEC 62321:2008 Ed.1 Sec.10	ICP-OES	2 mg/kg
Cadmium(Cd)	IEC 62321:2008 Ed.1 Sec.10	ICP-OES	2 mg/kg
Mercury(Hg)	IEC 62321:2008 Ed.1 Sec.7	ICP-OES	2 mg/kg
Hexavalent Chromium(Cr(VI))	IEC 62321:2008 Ed.1 Annex C	UV-Vis	2 mg/kg
Chlorine(Cl)	Refer to BS EN 14582:2007	IC	10 mg/kg
Bromine(Br)	Refer to BS EN 14582:2007	IC	10 mg/kg

Test Result(s)

Tested Item(s)	Content
Lead(Pb)	N.D.
Cadmium (Cd)	N.D.
Mercury(Hg)	N.D.
Hexavalent Chromium(Cr(VI))	N.D.
Tested Item(s)	Content
Halogen(s)	
Chlorine (Cl)	N.D.
Bromine (Br)	N.D.

Note: The sample had been dissolved totally tested for Lead, Cadmium, Mercury.

-MDL = Method Detection Limit

-N.D. = Not Detected (<MDL)

-mg/kg = ppm = parts per million


Test Report

Report No. RLSZD000842980004


Page 3 of 4

Test Process


1. Lead(Pb), Cadmium(Cd)




2. Mercury(Hg)



3. Hexavalent Chromium(Cr(VI))



4. Chlorine(Cl), Bromine(Br)




Test Report

Report No. RLSZD000842980004

Page 4 of 4

Photo(s) of the sample(s)



*** End of report ***

This report is considered invalidated without the Special Seal for Inspection of the CTI. This report shall not be altered, increased or deleted. The results shown in this test report refer only to the sample(s) tested. Without written approval of CTI, this test report shall not be copied except in full and published as advertisement.

Building C, Hongwei Industrial Zone, Baoan 70 District, Shenzhen